
## Storage & Usage Conditions for the Securetec DrugWipe®

Using DrugWipe® after improper storage may lead to false or inaccurate results. Improper storage at high temperatures or frost, will damage the test.

Below is a guide on what temperatures are acceptable to storage DrugWipes® and their usage conditions.



- Avoid temperatures above 40 °C! Test performance will be affected.
- > 5 40 °C DrugWipe® can be used in the temperature range of 5 to 40 °C.
- → 5 25 °C Store DrugWipe® in the long-term at temperatures from 5 to 25 °C.

DrugWipe® use below 5 °C, allow the test to reach room temperature prior to use (e.g. warm-up in your hands).

Freezing damages the tests. Even transport below 0 °C may cause breakage of the glass ampoule.



<sup>\*</sup>It is also important to note that DrugWipe® should only be used within its expiry range.